



Specification

1.00869.0500 Entellan™ new for cover slipper for microscopy

Specification

Refractive index (n 20/D)	1.490 - 1.500	
Fluorescence (as quinine at 365 nm)	≤ 250	ppb
Viscosity (20 °C)	500 - 600	mPa*s
Suitability for microscopy	passes test	

Dr. Timm Schlegelmilch

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.